# **ULT90**ULTRA LOW FREEZER VALIDATION SYSTEM



#### **Features**

- NIST Traceable Calibration Certificates
- FDA 21 CFR Part 11 Compliant
- IQ/OQ/PQ Validation Protocols
- Splash Proof Cryo-Temp Data Logger
- -86 °C to +35 °C Temperature Range
- 0.1 °C Resolution and ±1.0 °C Accuracy
- 1 Year Battery Life at -80 °C
- User-Programmable Temperature Alarms with LED's (OK, Warn, and Alarm)
- Hardware and Software Password
   Protection
- Simultaneous Starts, Memory Wrap
- User-Programmable Start Times,
   Sample Intervals, Logger
   Descriptions, Unique ID's

#### **Applications**

- Ultralow Freezer Validation
- Ultralow Freezer Mapping
- Ultralow Freezer Stabilization Studies
- Temperature Surveys
  - Reveal stratification of air
  - Reveal hot spots
  - Reveal anomalies in chilling systems




Available with 9, 15 or a custom number of Cryo-Temps

Ultra Low Temperature Freezer Validation System can measure and log data as low as -86 °C. It can be placed in small and large freezers of varying configurations.

The system provides a complete stand-alone temperature mapping and validation solution for freezers to ensure uniform storage and testing conditions. The ULT90 includes a USB docking station with 6 feet of cable (IFC300), the Cryo-Temp data logger, and MadgeTech's FDA 21 CFR Part 11 Secure Software and IQ/OQ/PQ validation protocols. For larger freezers, loggers can be plotted in one graph to show variances in different parts of the freezer. Finally, as failure to monitor external freezer temperatures may invalidate internal freezer temperatures, an additional Cryo-Temp can be included for placement outside of the freezer.

#### **DATA LOGGER SOFTWARE**



#### **Software Features:**

- Multiple graph overlay
- Statistics
- Digital calibration
- Zoom in/ zoom out
- Lethality equations (F0, PU)
- Mean Kinetic Temperature
- Full time zone support
- Data annotation
- Min./Max./Average lines
- Data table view
- Automatic report generation
- Summary view
- Multilingual



## **ULT90 SPECIFICATIONS\* (CRYO-TEMP DATA LOGGER)**

	NTC Thermistor -86 °C to +35 °C recommended,	Password Protection:	An optional password may be programmed into the device to restrict access to configuration options. Data may be read out without the password.
Measurement Range:	may be used up to +55 °C	Calibration:	Digital calibration is available to the user through
Resolution:	0.1 °C	Calibration:	software
Calibrated Accuracy:	±1.0 °C	Calibration Date:	Automatically recorded within device
Specified Accuracy Range:		*Battery Type:	3.6V lithium battery (LTC-7PN), *Must be sent to MadgeTech for battery replacement.
Memory:	32,767 readings; software configurable memory wrap		2 years typical @ 25 °C, 1 min. reading intervals
Reading Rate:	5 seconds to 30 minutes, selectable in	Battery Life:	1 year typical @ –80 °C, 1 min. reading intervals
Start Modes:	Software Software programmable immediate start, delay start up to 30 days in advance or manual start with magnetic switch.	Time Accuracy:	$\pm 2$ minutes/month max. at 20 °C to 30 °C $\pm 5$ minutes/month max at –80 °C
		Data Format:	Date and time stamped °C, °F, K, °R
Temperature Alarm:	Software programmable high and low alarm limits and high and low	Computer Interface:	USB docking station to PC (IFC300) 38,400 baud
		Software:	XP SP3/Vista/Windows 7 Not Compatible with 64 bit systems
	<ul> <li>Each setpoint may be individually enabled or disabled</li> </ul>	Operating Environment:	-86 °C to +35 °C (may be used up to +55 °C); 0 to 100% RH
Alarm Delay:	Software programmable time delay for warn/alarm limits. Logger will not		
		Dimensions:	3.3" x 2.1" x 0.4" (83 mm x 54 mm x 11 mm)
		Ingress Protection/IP Rating:	IP64 (splash resistant, not submersible)
Status Indicators:	Green: Indicates safe conditions during logging	Enclosure:	ABS plastic, white, polycarbonate labeling
	Yellow: Indicates warn limits have been exceeded	Weight:	1.4 oz (40 g) each Total for packages: ULT90-9 - 50.9 oz (1443 g) ULT90-15 - 59.3 oz (1681 g)
	<ul> <li>Red: Indicate alarm criteria (limits/ delay) have been exceeded</li> </ul>	Compliance:	CE

BATTERY WARNING: FIRE, EXPLOSION, AND SEVERE BURN HAZARD. DO NOT RECHARGE, DISASSEMBLE, HEAT ABOVE 212 °F, INCINERATE OR EXPOSE CONTENTS TO WATER.

### **ORDERING INFORMATION**

MODEL		DESCRIPTION
	ULT90-9	Includes 9 Cryo-Temp data loggers with NIST Traceable calibration certificates, IFC300 docking station, MadgeTech's FDA 21 CFR Part 11 Secure Software and IQ/OQ/PQ validation protocols.
	ULT90-15	Includes 15 Cryo-Temp data loggers with NIST Traceable calibration certificates, IFC300 docking station, MadgeTech's FDA 21 CFR Part 11 Secure Software and IQ/OQ/PQ validation protocols.

Temperature Humidity **ASK ABOUT** Pressure **OUR OTHER** рΗ DATA Level LOGGERS Shock LCD Display Pulse/Event/State Current Voltage Wireless Intrinsically Safe Spectral Vibration Motion























